



TRANSPARENCY INTERNATIONAL-PAKISTAN

5-C, 2nd Floor, Khayaban-e-Ittehad, Phase VII,
Defence Housing Authority, Karachi.
Tel: (92-21)-35390408, 35390409, Fax: 35390410
E-mail: ti.pakistan@gmail.com
Website: www.transparency.org.pk

18th January, 2017

TL17/1801/7A

Hospital Director,
Hayatabad Medical Complex,
Peshawar.

Sub: Violation Khyber Pakhtunkhwa Public Procurement Regulatory Authority Act 2012,
Generalize Specifications by Mentioning Made: England/ EU/Japan/ USA. In HMC
Tender Peshawar Dated: 20.12.2016.

Dear Sir,

Transparency International Pakistan has received a complaint from M/s AQK Traders, Peshawar against Hayatabad Medical Complex in procurement of 'ICU ventilator with Air Compressor' for stating country origin: England/ EU/Japan/ USA (Copy Encl.).

The complainant has alleged that Hayatabad Medical Complex Peshawar, in tender documents for purchase of ventilator has restricted the bidders to supply ventilator from England/ EU/Japan/ USA origin only.


In case procuring agency has to mentioned any brand name or country of origin it should by adding "**or equivalent**" with it. Hence tender is in violation under KPK PPRA Act 2012 Section No. 23 (4). Stated as under;

23. Bid Solicitation documents --- (4) Each procuring entity shall solicit bids based on performance or functional specifications and not on restrictive or proprietary specifications of a particular brand.

The above information is forwarded for the purpose of avoiding mis-procurement charge under Rule No 54, with request to re-invite the tender under the prescribed procedures or issue a corrigendum and extend date accordingly.

Transparency International Pakistan is striving for across the board application of Rule of Law, which is the only way to stop corruption.

With Regards,


Sohail Muzaffar
Chairman

Copies forwarded for the information with request to take action under their mandate to:

1. Chief Minister, Khyber Pakhtunkhwa.
2. Director General, NAB, Khyber Pakhtunkhwa.
3. Chief Secretary, Khyber Pakhtunkhwa.
4. Registrar, Peshawar High Court.
5. Managing Director, KPPRA, Peshawar.

ALS IN SURGICAL & MEDICAL EQUIPMENT AND GENERAL ORDER SUPPLIERS
ON APPROVED LIST OF HEALTH DEPARTMENT a q k TRADERS

Muslim Commercial Bank

Bank A/c No: 0504193171000461

Branch: Haji Khan Building Peshwar Income Tax No3552876-1

Bank Code: 0278

Sales Tax No : 2100355287611

Vender No : 30124495



S.No: AQK/110A
Dated: 13-Jan-2017

To
**Transparency International
Pakistan.**

**Subject: Violation of Generalize Specification by mentioning made:
England/EU/Japan/USA. In HMC Tender Peshawar Dated: 20.12.2016.**

Sir,

With reference to tender advertized by Hospital Director Hayatabad Medical Complex Peshawar for Purchase of Ventilator dated 20/12/2016 in News Paper "SUBH" Peshawar, it is observed that the Department has mentioned specific Made such as **England/EU/Japan/USA** in required specification which eliminate many participant from submitting the tender, thus open competition cannot be made. As we are the authorize distributor of same required product of "MER" an International Brand with "FDA & CE" approved (Made In Korea) products already satisfactorily supplied & installed in major Government Institutions of KPK.

- a. DHQ Hospitals KPK (References can be provide)
- b. Teaching Hospitals KPK (References can be provide)

Kindly look into this matter to make it generalize and open for every Company. News paper cutting and specifications are attached for your reference.

Thanks & Regards

Director
Abdul Qadir Khan Khel

AQK Traders
Peshawar

CC: Hospital Director HMC Peshawar



حیات آباد میڈیکل کمپلیکس پشاور

ٹینڈر نوٹس

Under National competitive Bidding (NCB) For the year 2016-17

ہسپتال ہذا کو سال 2016-17 کے لیے بہترین قیمت رکھنے والی قومی اور بین الاقوامی تیار کنندگان اور کمپنی کے مجاز اجینٹس سے مندرجہ ذیل آلات اور دیگر اشیاء کیلئے For & C&F Single Stage One & Two - envelope Bidding (Financial Bid) کے تحت چھپدی گئی سیریل نمبرز کے مطابق ٹینڈرز درکار ہیں۔ پیشکش کے ساتھ ہر آئٹم کے سامنے دی گئی رقم کا کال لیٹرز (زیر تحفظ کی نام) منسلک ہونا ضروری ہے۔

S.No	Name of Items	CDR Register	CDR Un-Register
Single Stage one Envelope rule 6 (2a)			
1	3rd Time ENT Disposable.	(20000)	(50000)
2	2nd time Office Furniture	(100000)	(250000)
Single Stage two Envelope Rule 6 (2b)			
3	Purchase of Radiology Equipments	2%	5%
4	Purchase of Neuro Surgery Equipment	2%	5%
5	Purchase of Peads Equipments	2%	5%

(1) تمام منگور شدہ نرخ 30/06/2017 تک قابل عمل رہیں گے۔ (2) ٹینڈر ایڈنگ Documents سنٹرل پروکیورمنٹ اینڈ پراجیکٹس، حیات آباد میڈیکل کمپلیکس سے ذمہ داری کے ساتھ حاصل کیے جاسکتے ہیں۔ (3) صرف وہ ٹینڈرز قابل قبول ہوں گے جو کاپی کے لیے ٹینڈر پر مابین شدہ ہوں۔ ہاتھ سے لکھے ہوئے ٹینڈرز قابل قبول نہ ہوں گے۔ (4) ٹینڈر دینے والے کو 10:30 AM 03-01-2017 بوقت 10:30 AM پہلے ذمہ داری کو بذریعہ رجسٹرڈ ڈاک یا کوریئر سروس مل جانی چاہئے دیر سے پہنچنے والے ٹینڈرز وصول نہیں کیے جائیں گے۔ (5) ٹینڈرز ہندوستان کے رجسٹرڈ بینکاروں کے ذریعہ دیئے جائیں گے۔ (6) جو کمپنیاں DGHS KPK کے ساتھ رجسٹرڈ اپری کو ایلیٹائیڈ ہوں وہ مندرجہ بالا (2%) مطلوب کال لیٹرز جمع کروائیں۔ لوہو و کمپنیاں DGHS KPK کے ساتھ رجسٹرڈ اپری کو ایلیٹائیڈ نہ ہوں وہ (5%) کے برابر کال لیٹرز جمع کروائیں۔ (7) Proprietary Certificate درکار ہوا تو ٹینڈر کے ساتھ منسلک کرنا لازم ہے۔ (8) تمام کمپنیوں کو رجسٹرڈ کی پالیسی کے مطابق نافذ العمل ہوں گے۔ (9) Active Tax Payer List پر ہونا لازم ہے۔ (10) حکومت کیساتھ مقدمات ہوں وہ کوئی ٹینڈر نہیں۔ (11) کامیاب ٹینڈرز ہندوستان کی آمدی آمدی 90 دن کے اندر لوہو کو مل چلائی 30 دن کے اندر اندر چلائی کرنے کا پابند ہوگا ورنہ گورنمنٹ قوانین کے مطابق مذکورہ چلائی آرڈر پر جرمانہ عائد کر دیا جائے گا۔ (12) شرائط ٹینڈرز قابل قبول نہ ہوں گے۔ (13) صرف ایجوکیشن کیلئے ٹیکنیکل اور فنانسنگل کویشن ایکٹ الگ ہونی چاہئے۔ (14) تمام کمپنیوں کو رجسٹرڈ کی پالیسی کے مطابق نافذ العمل ہوں گے۔ نوٹ: زیر تحفظ کی کوئی ٹینڈر مرہبہ قانون (KPPRA) Procurement Rules 2014 سب سے 47 منگور کرنے کا حق حاصل ہے۔ ٹینڈر جمع کرانے کا پتہ: حیات آباد میڈیکل کمپلیکس، پشاور

فون: 091-9217189-47, 091-9217140-091 www.hmckp.gov.pk

Say no to corruption

INF (P) 7552

ہاسپتال ڈائریکٹر حیات آباد میڈیکل کمپلیکس پشاور

لیجسلیٹو اور 16-12-20

Trigger sensitivity indication

Mounted on trolley with lockable wheel

Alarms:

Apnea

AC power failure

High and low expired minute volume

High and low peak air way pressure

High and low breathe rate

High and low CO₂

PiO₂ variation

Gas supply source failure

PEEP loss

Low battery

Nebulizer:

Built in nebulizer for nebulization of the patient during ventilation

Power supply Requirements

220 V 50 Hz with internal chargeable battery backup min for 1 hr.

Autoclaveable reusable patient circuit for adult, paed & Neonates (1 each)

Compressed Air Supply

The ventilator should be driven on central air and oxygen system and have compressor as back up source or having built in air turbine system.

Humidifier:

Automatic compensation (Servo) controlled heated humidifier with temperature monitoring at air way and humidification chamber with alarm for low/ high limits with water trap in the patient circuit.

Scope of Supply:

Ventilators on trolley with wheels and brake facility

Bacterial filter (05 nos.)

Flow Sensors (05 nos.)

Oxygen Sensors (05 nos.)

Software backup should be provided

Breathing circuit all ranges (disposable, qty.05 each for Adult, Peads, Neonates.)

Air Filter

Oxygen and air connections/probes

Training CD/DVD

05 years comprehensive warranty with parts

Service Manual and operator manual must be provided

Quality Certificate: CE / FDA (510K) / JIS (MHLW)

Country of Origin: England, EU, USA, Japan.

Quarterly Periodic Preventive Maintenance shall be done by the company in the presence of Hospital Biomedical Engineer till the warranty period.

Three Weeks Technical Training shall be provided to the Biomedical Engineering Department from the manufacturer by the vendor/supplier.

